



Declaration of Conformity

ESCORT Temperature recorders iLog EI-IN-(X)-32-L, EI-1E-(XX)-32-L, EI-2E-(XX)-32-L and iLog Humidity EI-HS-D-32-L meet the guidelines in accordance with the Electromagnetic Compatibility (EMC) Directive 89/336/EEC

The products have been tested for conformance with CE mark requirements and have passed in accordance to the following standards:

EN 55022: 1998 Limits and methods of measurement of radio disturbance of information technology equipment.

Report No.: 30708.1 – 30709.1
Report Date: 29 July 2003
(IANZ) Accredited Lab: EMC Technologies (NZ) Ltd

EN 50082-1: 1997 Electromagnetic compatibility: Generic immunity standard. Part 1. Residential, commercial and light industry. (61000-6:2001)

Report No.: 30708.2 – 30709.2
Report Date: 29 July 2003
(IANZ) Accredited Lab: EMC Technologies (NZ) Ltd

Christopher J. Smith
Site/Quality Management Systems
Manager

(X) or (XX) stands for Temperature Range.
Technical information available upon request

UNCONTROLLED when printed. Check the MASTER LIST to verify this is the correct revision before use.

Authorized QMS Document for Cryopak Verification Technologies. Reference: ISO 9001:2008.

ENF-005 Revision 2. Page 1 of 1 Produced by ISO Administrator
